

The Regional Brewer Calibration Center - Europe: an overview of the X Brewer Intercomparison Campaign

Alberto Redondas^{1*}, Juan J Rodríguez-Franco¹, Javier López-Solano^{1,2}, Virgilio Carreño¹, Sergio Fabián León-Luis¹, Bentorey Hernández-Cruz^{1,2}, Alberto Berjón^{1,2}

Abstract

The X Regional Brewer Calibration Center - Europe (RBCC-E) intercomparison campaign was held at El Arenosillo atmospheric sounding station of the Instituto Nacional de Técnica Aeroespacial (INTA) during the period of May 25th – June 5th, 2015. This X Brewer intercomparison campaign was a joint effort of EUBREWNET and the Area of Instrumentation and Atmospheric Research of INTA, with the support of COST Action 1207. Twenty one Brewer instruments from eleven countries participated at the El Arenosillo campaign. In addition, the solar UV irradiance calibration was performed by the traveling reference standard QASUME of the World Calibration Center for UV (WCC-UV).

This work shows a general overview of the ozone comparison focused on the correction of the stray light effect for the single-monochromator Brewer spectrophotometer derived by the comparison with a reference double-monochromator Brewer. At the beginning of the campaign, 16 out of the 21 participating Brewer instruments (76%) agreed to better than $\pm 1\%$, and 10 instruments (50%) agreed to better than $\pm 0.5\%$. After applying the final calibration that included the stray light correction, all working instruments agreed at the $\pm 0.5\%$ level.

¹ Regional Brewer Calibration Center for Europe, Izaña Atmospheric Research Center, Agencia Estatal de Meteorología, Tenerife, Spain

² Departamento de Ingeniería Industrial, Universidad de La Laguna, Tenerife, Spain

*Corresponding author: aredonasm@aemet.es

Contents

Introduction	1
1 The Brewer spectrometer calibration	2
1.1 ETC transfer and stray light	3
2 Intercomparison Results	4
2.1 Reference Calibration	4
2.2 Blind Days	5
2.3 Final comparison	6
Conclusions	7
Acknowledgments	8
References	8

Introduction

In November 2003 the WMO/GAW Regional Calibration Center for RA-VI region (RBCC-E) was established at the Observatory Izaña of AEMET, Canary Islands (IZO). RBCC-E owns a full calibration and reference-maintenance equipment composed by three Brewer spectroradiometers (the IZO Triad): a Regional Primary Reference (Brewer #157), a Regional Secondary Reference (Brewer #183) and a Regional Traveling Reference (Brewer #185) which can be transported for calibration campaigns outside IZO. The Regional Brewer calibration Center transfers the calibration from the World Reference

Triad in Toronto. Due to uncertainties on the future maintenance of the World Triad, in the 2011 meeting the WMO scientific advisory group (WMO-SAG) authorized the RBCC-E to transfer its own calibration obtained by the Langley method.

RBCC-E regular inter-comparisons are held annually, alternating between Arosa in Switzerland and the El Arenosillo sounding station of the Instituto Nacional de Técnica Aeroespacial (INTA) at Huelva in the south of Spain. Since 2005, a total of 130 Brewer ozone spectrophotometer calibrations have been performed in these calibration campaigns (see the campaign reports at the RBCC-E website [1] and the GAW reports of the VII [2], VIII [3], and IX [4] intercomparison campaigns). In addition to the regular intercomparisons, the RBCC-E performs two types of campaigns supported by the ESA CalVal project: the NORDIC campaigns, with the objective to study the ozone measurements at high latitudes, and the Absolute calibration campaigns performed at IZO with the participation of Brewer and Dobson reference instruments (Figure 1).

The X RBCC-E campaign

From May 25th to June 5th, 21 Brewer spectrophotometers from 11 countries (see Table 1) took part in the X RBCC-E campaign held at the El Arenosillo atmospheric sounding station (Huelva, Spain). Besides the ozone calibration, a solar UV irradiance calibration was performed by the traveling reference standard QASUME instrument of the World Cali-




Figure 1. Instruments calibrated since 2003 by the RBCC-E in regular campaigns (at Huelva and Arosa), Nordic intercomparisons, and the Absolute calibrations performed at the Izaña Observatory.

bration Center for UV (WCC-UV). The X RBCC-E campaign was the result of the collaboration between COST Action 1207 “EUBREWNET” [5] and the Area of Instrumentation and Atmospheric Research of INTA.

The aim of COST Action 1207 “EUBREWNET” is to establish a coherent network of European stations equipped with Brewer spectrophotometers for the monitoring of total ozone, spectral UV radiation and aerosol optical depth in the UV spectral range, ensuring sustainable operation in the long-term. Among the primary aims of EUBREWNET is to harmonize operations and develop approaches, practices and protocols to achieve consistency in quality control, quality assurance and coordinated operations, as well as to eliminate duplication of efforts at individual stations to achieve separately best practice and accuracy. It also aims at establishing knowledge exchange and training, and at opening up a route to link with international agencies and other networks globally. Around 50 Brewer spectrophotometers are deployed in Europe, independently funded by national institutions.

In parallel to the campaign, COST Action 1207 organized several experiments to advance the quality of the Brewer data. Among them, studies on the Dead Time determination using the direct-sun measurements, characterization of the temperature dependence of the Brewer diffuser, cosine response measurements, investigation of the effects of polarization of the input window, aerosol optical depth calibration, stray light characterization, comparison of total ozone between Phaethon and Brewer instruments, and an intercomparison of UV reference lamps.

1. The Brewer spectrometer calibration

The Brewer instrument measures the intensity of direct sunlight at six wavelengths (λ) in the UV (303.2, 306.3, 310.1, 313.5, 316.8, and 320.1 nm) each covering a bandwidth of 0.5 nm (resolution power $\lambda/\delta\lambda$ of around 600). The spectral measurement is achieved by a holographic grating in combina-

Table 1. Participating instruments during the X RBCC-E campaign.

No.	Country	Brewer ID	Participants
1	Greece	005	Alkis Bais
2	Canada	017	Kenneth Clark Lamb
3	Spain	033	Juan R. Moreta
4	Russian Federation	044	Vadim Shirotov
5	Spain	070	Juan R. Moreta
6	UK	075	John Rimmer
7	Spain	117	Juan R. Moreta
8	UK	126	John Rimmer
9	Spain	150	J. M. Vilaplana
10	Spain	151	Juan R. Moreta
11	Netherlands	158	Oleksii Marianenko
12	Switzerland	163	Julian Groebner
13	Spain	166	Juan R. Moreta
14	UK	172	John Rimmer
15	Spain	185	Alberto Redondas
16	Spain	186	Juan R. Moreta
17	Algeria	201	Boukela Lamine
18	Denmark	202	Paul Eriksen
19	Finland	212	Tomi Karprinen
20	Denmark	228	Niss Jepsen
21	Netherlands	230	Keith M. Wilson

tion with a slit mask which selects the channel to be analyzed by a photomultiplier. The longest four wavelengths are used for the ozone calculation. Based on the Lambert-Beer law, the total ozone column in the Brewer algorithm can be expressed as:

$$X = \frac{F - ETC}{\alpha\mu} \quad (1)$$

where F are the measured double ratios corrected for Rayleigh effects, α is the ozone absorption coefficient, μ is the ozone air mass factor, and ETC is the extra-terrestrial constant. The F , α and ETC parameters are weighted functions at the operational wavelengths with weighting coefficients w :

$$F = \sum_i^4 w_i F_i - \frac{p}{p_0} \beta_i \mu \quad (2)$$

$$\alpha = \sum_i^4 w_i \alpha_i \quad (3)$$

$$ETC = \sum_i^4 w_i F_{0i} \quad (4)$$

where, β_i are the Rayleigh coefficients, p is the climatological pressure at the measurement site, p_0 is the pressure at sea



Figure 2. Panoramic view of the 21 Brewer spectrophotometers on the terrace of the El Arenosillo atmospheric sounding station, Huelva, coming from Canada (1), Netherlands (2), United Kingdom (3), Switzerland (1), Finland (1), Greece (1), Denmark (2), Russia (1), Algeria (1) and Spain(7).

level, and F_0 are the individual extra-terrestrial constants at each wavelength. The weights $w = [1, -0.5, -2.2, 1.7]$ have been chosen so as to minimize the influence of SO_2 and verify:

$$\sum_i^4 w_i = 0 \quad (5)$$

$$\sum_i^4 w_i \lambda_i = 0 \quad (6)$$

This widely eliminates absorption features which depend, in local approximation, linearly on the wavelength, like for example the contribution from aerosols.

We can divide the calibration in three steps:

1. Instrumental, wavelength, and ETC transfer: the Instrumental calibration includes all the parameters that affect the measured counts (F), in particular Dead Time correction, Temperature coefficients and Filter attenuation.
2. Wavelength calibration to determine the ozone absorption coefficient: the so-called “dispersion test” are used to obtain the particular wavelength for the instrument and the slit, or instrumental function, of each spectrophotometer. Note that the precise wavelengths of every Brewer spectrophotometer are slightly different from instrument to instrument.
3. Finally, the ETC transfer is performed by comparison with the reference or, in the case of the reference instruments, by the Langley method.

The calibration process can be considered as cycle changes. Instrumental and/or wavelength calibration will affect the final ETC and changes in the wavelength calibration will affect also to the final ETC. For this reason the calibration campaigns are scheduled in three different periods:

1. **Blind days:** the first days of the campaign are dedicated to determine the current status of the instrument, during this period modifications of the instrument are not allowed.
2. **Characterization:** after the determination of how the instrument is measuring, the next days are dedicated to characterize the instrument and perform the necessary adjustments and maintenance. The instrumental and wavelength calibration must be finished at the end of this period.
3. **Final days:** the period where the ETC transfer is performed when the instrument is fully characterized and stable.

1.1 ETC transfer and stray light

The transfer of the calibration scale (namely ETC) is done side by side with the reference instrument. Once we have collected enough near-simultaneous direct sun ozone measurements, we calculate the new extraterrestrial constant after imposing the condition that the measured ozone will be the same for simultaneous measurements. In terms of Equation 1, this leads to the following condition:

$$ETC_i = F_i - X_i^{reference} \alpha \mu \quad (7)$$




Figure 3. Distribution of individual ETC values determined by simultaneous measurements. In the x axis, the ozone slant column is shown divided by 1000 (OSC are expressed in cm). In this particular Brewer, the stray light is clearly shown at values above 0.6 for the scaled ozone slant column.

For a good characterized instrument, the ETC determined values show a Gaussian distribution and the mean value is used as the instrument constant. One exception to this rule is the single monochromator Brewer models (MK-II and MK-IV) which are affected by stray light. In this case, the ETC distribution shows (see Fig. 3) a tail at the lower ETC values for high Ozone Slant Column (OSC, the product of the total ozone content by the airmass). For this type of Brewer, only the stray-light free region is used to determine the ETC, generally from 300 to 900 DU OSC, depending on the instrument.

The stray light effect can be corrected if the calibration is performed against a double monochromator instrument, assuming that it can be characterized following a power law of the ozone slant column

$$F = F_o + k(X\mu)^s \quad (8)$$

where F are the true counts and F_o the measured ones.

$$ETC_i = ETC_o + k(X\mu)^s \quad (9)$$

where ETC_o is the ETC for the stray light free OSC region and k and s are retrieved from the reference comparison (Figure 4). These parameters, determined in several campaigns, have been found to be stable and independent of the ozone calibration.

As the counts (F) from the single brewer are affected by stray light, the ozone is calculated using an iterative process:

$$X_{i+1} = X_i + \frac{k(X_i\mu)^s}{\alpha\mu} \quad (10)$$

Only one iteration is needed for the conditions of the intercomparison, up to 1500 DU. For ozone slant path measurements in the 1500–2000 DU range, two iterations are




Figure 4. The stray light parameters k and s are determined by a nonlinear fit using the ETC determined from the stray-light free region as first guess parameters. The red horizontal line indicate the ETC constant retrieved from the fit whereas the green line shows the initial guess .

enough to correct the ozone (Figure 5). These stray light corrections are now implemented in the standard processing of EUBREWNET.




Figure 5. Percentage differences in ozone with respect to the reference vs. Ozone Slant Path. In blue, using the final configuration constants, and in black and red after the stray light correction has been applied, with one and two iterations, respectively. Data are averaged in $\pm 50DU$ intervals, the shadow area represent the one standard deviation.

2. Intercomparison Results

2.1 Reference Calibration

The RBCC-E triad is regularly calibrated, performing the instrumental characterization and wavelength calibration monthly. The three instruments are independently Langley Calibrated following the methodology described in Ref. [6]. Before and

after the intercomparison campaigns, the traveling instrument is compared with the two static instruments to verify that the calibration do not change during transport (Figures 6 and 7).




Figure 6. Box plot of the ozone percentage deviation from the mean before the X RBCC-E campaign at Huelva in 2015.




Figure 7. Box plot of the ozone percentage deviation from the mean after the X RBCC-E campaign at Huelva in 2015.

The campaign is a good opportunity to compare reference instruments, that is instruments that are used to transfer calibrations. In this campaign participated Brewer #017, managed by International Ozone Services (IOS) and directly calibrated to the Environmental Canada Toronto Triad, and Brewer #158 managed by Kipp & Zonen, manufacturer of the Brewer spectrophotometer. The agreement between the references is quite good, with differences lower than 0.5% for OSC lower than 900 DU¹(see Table 2).

¹Note that Brewer #017 is a single-monochromator instrument and is affected by the stray light underestimating the ozone at high OSC (OSC > 600 DU)




Figure 8. Comparison of reference instruments during the X RBCC-E campaign: Relative differences to the reference using the initial configuration during the campaign, in red for the IOS Brewer #017, stars for the original observations and cross for stray light corrected, the K&Z Brewer #158 in blue squares. The gray points are the relative differences to the reference for all participating instruments

2.2 Blind Days

A blind comparison with the standard Brewer is performed at the beginning of the campaign, this exercise gives us an idea of the initial status of the instrument, i.e. how well the instrument performed using the original calibration constants (those operational at the instrument's station). Possible changes of the instrument response due to the travel can be detected through the analysis of internal tests performed before and after the travel.




Figure 9. Percentage mean difference for the simultaneous direct sun measurements with the reference for all the participating instruments, with and without the standard lamp correction, on the stray-light free OSC region (OSC<900).

The instruments are working during this period with their home calibration and the ozone is calculated using these calibration constants. The Standard Lamp (SL) test is an ozone measurement using the internal halogen lamp as a source. This test is performed routinely to track the spectral response of the instrument and, therefore, the ozone calibration. A reference value for the SL, the so-called R6 ratio, is provided as

part of the calibration of the instrument. The ozone is routinely corrected assuming that deviations of the R6 value from the reference value are the same that changes in the Extraterrestrial constant (ETC). This then described by the Standard Lamp correction:

$$ETC_{new} = ETC_{old} - (SL_{ref} - SL_{measured}) \quad (11)$$

The analysis of the SL historic record is one of the principal tools to establish the stability of the instrument calibration. Moreover the comparison with a reference during calibration campaigns is the most suitable tool to determine if the observed R6 changes are related or not with changes in the ETC constant.




Figure 10. Standard lamp R6 difference with respect to the R6 reference value from the last calibration during the blind days, before the maintenance. Variations within the ± 10 units range ($\sim 1\%$ in ozone) are considered normal, whereas larger changes would require further analysis of the instrument performance.

During the Huelva 2015 intercomparison campaign most of the instruments agreed on average with the corresponding R6 reference value within ± 10 units, which is about 1% in ozone. Some instruments showed deviations of R6 values to the reference larger than 20 units (Fig. 10). The comparison with a standard instrument is the only way to assess whether

the SL correction properly tracks changes on the calibration constants or the changes observed are just due to changes of the lamp’s spectral emission (Fig. 9). In some instruments, for example Brewer #075, the SL correction improves the comparison, whereas for others like #165 the opposite happens. This will determine if a re-evaluation of the ozone observations during calibrations are required after an analysis of the history of the instrument.

Table 2. Summary of mean percentage difference before calibration, without and with Standard Lamp Correction, and after the calibration, on the last column with the stray light correction applied.

Brewer ID	No SL corr.	SL corr.	Blind	Final	Stray Light corr.
005	-	-	-1.93	-0.2	-0.08
017	-0.31	-0.49	-0.98	-0.95	0.11
033	-0.8	-1.77	-1.09	-1.83	-0.48
044	-2.04	0.13	-0.21	-0.27	0.2
070	-0.73	-0.42	-0.71	-0.53	0.18
075	-3.42	-0.71	-1.2	-0.8	-0.2
117	-3.38	-0.45	-0.68	-0.6	0.04
126	-1.25	-1.41	-1.36	-0.29	-0.08
150	-0.45	-1.07	-0.45	-0.27	-
151	-17.36	9.94	7.95	0.67	0.83
158	-0.54	-2.45	-0.54	0.05	-
163	-1.5	-4.16	-1.5	-0.06	-
166	-0.15	1.45	-0.24	-0.58	-
172	-0.67	-0.67	-0.67	-0.01	-
186	0.13	-0.34	0.13	-0.05	-
201	1.21	0.52	0.52	0.09	-
202	-1.39	-0.95	-0.95	-0.06	-
214	1.42	1.19	1.19	-0.01	-
228	-1.93	-0.4	-0.4	-0.1	-
230	-0.15	-3.48	-0.15	0.36	-

Table 2 shows the percentage mean difference for the simultaneous direct sun measurements with the reference for all the participating instruments, with and without the standard lamp correction, on the stray light free OSC region. With the exception of Brewer #151, that can not be considered an operative instrument, the maximum difference found is 1.5%. This is a really good result considering that most of the instruments were calibrated two years ago. The third column of the table shows the average of the best result for all the observation OSC range. This result is an estimation of the calibration agreement of the EUBREWNET network, with half of the instruments showing a perfect agreement within $\pm 0.5\%$, and 75% on the $\pm 1\%$ level.

2.3 Final comparison

We define the final days as those available after the maintenance work has been finished for each participating instrument. These days are used to calculate the final calibration constants, so we tried not to manipulate the instruments during this period. Furthermore, the SL R6 value recorded during the final days is normally adopted as the new reference value. It is also expected that this parameter will not vary more than 5




Figure 11. Differences between the daily standard lamp R6 ratio and the proposed R6 reference value during the final days.

units during this period. We show in Figure 11 the differences between the daily standard lamp R6 ratio and the proposed R6 reference value during the final days. As expected, the recorded SL values did not vary more than 5 units during this period.

Deviations of ozone values for all the participating instruments with respect to the RBCC-E traveling standard Brewer #185 are shown in Fig. 12 and summarized in Table 2. We have recalculated the ozone measurements using the final calibration constants, an in the case of single Brewer instruments, with and without the stray light correction as described in Sec. 1. All Brewers were calibrated using the one parameter ETC transfer method, i.e., the ozone absorption coefficient was derived from the wavelength calibration (dispersion test) and only the ozone ETC constant was transferred from the reference instrument. The so-called “two parameters calibration method” [7], where the ozone absorption coefficient is also calculated from the reference, is also used as a quality indicator. For all the instruments both the one parameter and the two parameters ETC transfer methods agreed to each other within the limit of ± 5 units for ETC constants and $\pm 0.3\%$ for ozone absorption coefficients, which is an indication of the quality of the calibration provided.

We achieved a good agreement with the reference instrument Brewer #185 using the final calibration constants,(see Fig. 12 and Table 2. With the application of the stray light cor-




Figure 12. Final days percentage mean difference for the simultaneous direct sun measurements with the reference for all the participating instruments, in blue without the stray light correction and in red with the correction applied to single Brewer spectrophotometers.

rection to the single Brewer instruments, all of the instruments are within the $\pm 0.5\%$ agreement range.

Summary

Blind days Before the X RBCC-E campaign, using a two-years old calibration

- 16 Brewer spectrophotometers ($\sim 75\%$ of the participating instruments) are inside the 1% agreement range.
- 10 Brewer spectrophotometers ($\sim 50\%$) are inside the $\pm 0.5\%$ range, i.e., show a perfect agreement.
- The max average error was 1.5% for operational Brewer instruments within stray-light free conditions (OSC < 700 DU).

Final days After the new calibration was issued at the end of the X RBCC-E campaign,

- Large errors of up to 4% can be expected for single-monochromator Brewer instruments operating at OSC larger than 1000 DU.
- The implementation of the stray light correction in the calibration of single Brewer instruments improved their performance.
- All participating Brewer spectrophotometers are within the $\pm 0.5\%$ agreement range.

Acknowledgments

All this work would not be possible without all the Brewer operators who participate in the campaigns.

This article is based upon work from COST Action 1207 “EUBREWNET”, supported by COST (European Cooperation in Science and Technology). This work has been supported by the European Metrology Research Programme (EMRP) within the joint research project ENV59 “Traceability for atmospheric total column ozone” (ATMOZ). The EMRP is

jointly funded by the EMRP participating countries within EURAMET and the European Union. We also gratefully acknowledge further support by the Fundación General de la Universidad de La Laguna.

References

- [1] Website of the Regional Brewer Calibration Center - Europe, <http://rbcce.aemet.es>.
- [2] A. Redondas, J. Rodríguez-Franco, J. Gröbner, U. Köhler, and R. Stuebi. *Seventh Intercomparison Campaign of the Regional Brewer Calibration Center Europe (RBCC-E)*. Number 216 in WMO/GAW Reports. World Meteorological Organization, 2015.
- [3] A. Redondas and J. Rodríguez-Franco. *Eighth Intercomparison Campaign of the Regional Brewer Calibration Center Europe (RBCC-E)*. Number 223 in WMO/GAW Reports. World Meteorological Organization, 2015.
- [4] A. Redondas and J. Rodríguez-Franco. *Ninth Intercomparison Campaign of the Regional Brewer Calibration Center Europe (RBCC-E)*. Number 224 in WMO/GAW Reports. World Meteorological Organization, 2015.
- [5] Website of COST Action 1207 “EUBREWNET”, <http://www.eubrewnet.org/cost1207>.
- [6] A. Redondas. Ozone absolute langley calibration. *The Tenth Biennial WMO Consultation on Brewer Ozone and UV Spectrophotometer Operation, Calibration and Data Reporting*, Edited by C. T. McElroy and E. W. Hare, *Gaw Report. No. 176*, WMO TD No. 1420:12–14, 2007.
- [7] J. Staehelin, J. Kerr, R. Evans, and K. Vanicek. Comparison of total ozone measurements of Dobson and Brewer spectrophotometers and recommended transfer functions. *WMO TD*, (1147):20, 2003.